



# **Albumin (Plant)**

## **Microplate Assay Kit**

### **User Manual**

**Catalog # CAK1197**

(Version 1.2A)

Detection and Quantification of Albumin (Plant) Content in Tissue extracts, Powder Samples.

**For research use only. Not for diagnostic or therapeutic procedures.**

I. INTRODUCTION.....	2
II. KIT COMPONENTS.....	3
III. MATERIALS REQUIRED BUT NOT PROVIDED.....	3
IV. SAMPLE PREPARATION.....	4
V. ASSAY PROCEDURE.....	5
VI. CALCULATION.....	6
VII. TYPICAL DATA.....	7
VIII. TECHNICAL SUPPORT.....	7
IX. NOTES.....	7

## I. INTRODUCTION

Albumin (Plant) Microplate Assay Kit is a sensitive assay for determining albumin content in plant samples. The color intensity, measured at 595 nm, is proportionate to albumin content in the sample.

## II. KIT COMPONENTS

Component	Volume	Storage
96-Well Microplate	1 plate	
Assay Buffer	30 ml x 2	4 °C
Dye Reagent	20 ml x 1	4 °C
Standard	Powder x 1	-20 °C
Technical Manual	1 Manual	

### Note:

**Standard:** add 1 ml distilled water to dissolve before use, the concentration will be 2 mg/ml.

## III. MATERIALS REQUIRED BUT NOT PROVIDED

1. Microplate reader to read absorbance at 595 nm
2. Distilled water
3. Pipettor, multi-channel pipettor
4. Pipette tips
5. Mortar
6. Ice
7. Centrifuge
8. Timer
9. Lab rotator

#### **IV. SAMPLE PREPARATION**

##### **1. For tissue samples**

Weigh out 0.05 g tissue, homogenize with 0.5 ml Assay Buffer on ice, transfer it to centrifuge tube and mix on a lab rotator for 30 minutes; centrifuged at 10000g 4 °C for 10 minutes, take the supernatant into a new centrifuge tube and keep it on ice for detection.

##### **2. For powder samples**

Weigh out 0.05 g powder, add 0.5 ml Assay Buffer to dissolve, mix on a lab rotator for 30 minutes; centrifuged at 10000g 4 °C for 10 minutes, take the supernatant into a new centrifuge tube and keep it on ice for detection.

## V. ASSAY PROCEDURE

Add following reagents into the microplate:

Reagent	Sample	Standard	Blank
Sample	10 $\mu$ l	--	--
Standard	--	10 $\mu$ l	--
Distilled water	--	--	10 $\mu$ l
Dye Reagent	200 $\mu$ l	200 $\mu$ l	200 $\mu$ l
Mix, wait for 2 minutes, measured at 595 nm and record the absorbance.			

### Note:

- 1) Perform 2-fold serial dilutions of the top standards to make the standard curve.
- 2) The concentrations can vary over a wide range depending on the different samples.  
For unknown samples, we recommend doing a pilot experiment & testing several doses to ensure the readings are within the standard curve range.
- 3) Reagents must be added step by step, can not be mixed and added together.

## VI. CALCULATION

1. According to the weight of sample

$$\begin{aligned}\text{Albumin (mg/g)} &= (C_{\text{Standard}} \times V_{\text{Standard}}) \times (\text{OD}_{\text{Sample}} - \text{OD}_{\text{Blank}}) / (\text{OD}_{\text{Standard}} - \text{OD}_{\text{Blank}}) / \\ &\quad (V_{\text{Sample}} \times W / V_{\text{Assay}}) \\ &= (\text{OD}_{\text{Sample}} - \text{OD}_{\text{Blank}}) / (\text{OD}_{\text{Standard}} - \text{OD}_{\text{Blank}}) / W\end{aligned}$$

$C_{\text{Standard}}$ : the standard concentration, 2 mg/ml;

$V_{\text{Standard}}$ : the volume of standard, 10  $\mu\text{l}$  = 0.01 ml;


$V_{\text{Sample}}$ : the volume of sample, 10  $\mu\text{l}$  = 0.01 ml;

$W$ : the weight of sample, g;

$V_{\text{Assay}}$ : the volume of Assay Buffer, 0.5 ml.

## VII. TYPICAL DATA

The standard curve is for demonstration only. A standard curve must be run with each assay.



Detection Range: 0.02 mg/ml - 2 mg/ml

## VIII. TECHNICAL SUPPORT

For troubleshooting, information or assistance, please go online to [www.cohesionbio.com](http://www.cohesionbio.com) or contact us at [techsupport@cohesionbio.com](mailto:techsupport@cohesionbio.com)

## IX. NOTES